

LYRA E.T.K
DIGITAL DENTAL PROTOCOLS

Simple . Safe . Accessible


2024 GUIDA ALLA COMPATIBILITÀ DI IPHYSIO CON I VOSTRI T-BASE

INTRODUZIONE

iPhysio è stato progettato per adattarsi alla maggior parte dei sistemi implantari. Vi permette di migliorare la vostra pratica senza cambiare sistema. Questa guida mostra, per ogni marca di impianto, i t-base che è possibile utilizzare nella propria gamma.

Legenda di compatibilità :

- ✓ **Raccomandato** = il profilo di emergenza del t-base è identico a quello del riferimento iPhysio indicato. È indicato come tale nella libreria che forniamo al vostro laboratorio ("raccomandato").
- ✓ ~ **Altra scelta possibile** = il profilo di emergenza del t-base è compatibile con quello del riferimento iPhysio indicato. Compare con la dicitura "altro".
- X **Non compatibile** = il profilo di emergenza del t-base non è compatibile con quello del riferimento iPhysio indicato, è troppo differente. Per evitare errori di utilizzo, i t-base identificati come non compatibili non sono referenziati nella nostra biblioteca



STRAUMANN

	NC H3.5 / H5.5 GH1.0	NCH 3.5 / H5.5 GH2.0	NC H3.5 / H5.5 GH3.0	NC ASC	NC Bridge/bar Conical	NC Bridge/bar Cylindrical	NC TAN
SBLN_I.A1	~	X	X	✓	✓	✓	✓
SBLN_I.A2	✓	X	X	✓	✓	✓	✓
SBLN_I.A3	~	✓	X	X	X	X	✓
SBLN_I.A4	~	✓	X	X	X	X	✓
SBLN_I.B1	~	X	X	✓	✓	✓	✓
SBLN_I.B2	✓	X	X	✓	✓	✓	✓
SBLN_I.B3	~	✓	X	X	X	X	✓
SBLN_I.B4	~	✓	X	X	X	X	✓
SBLN_I.D1	~	X	X	✓	X	X	✓
SBLN_I.D2	✓	X	X	✓	X	X	✓
SBLN_I.D3	~	✓	X	X	X	X	✓
SBLN_I.D4	~	✓	X	X	X	X	✓

	RC H3.5 / H5.5 GH1.0	RC H 3.5 / H5.5 GH2.0	RC H3.5 / H5.5 GH3.0	RC ASC	RC Bridge/bar Conical	RC Bridge/bar Cylindrical	RC TAN
SBLR_I.B1	✓	X	X	✓	✓	✓	✓
SBLR_I.B2	✓	X	X	✓	✓	✓	✓
SBLR_I.B3	X	✓	X	X	X	X	✓
SBLR_I.B4	X	✓	~	X	X	X	✓
SBLR_I.C1	✓	X	X	✓	✓	✓	✓
SBLR_I.C2	✓	X	X	✓	✓	✓	✓
SBLR_I.C3	X	✓	X	X	X	X	✓
SBLR_I.C4	X	✓	~	X	X	X	✓
SBLR_I.D1	✓	X	X	X	~	~	✓
SBLR_I.D2	✓	X	X	X	~	~	✓
SBLR_I.D3	X	✓	X	X	X	X	✓
SBLR_I.D4	X	✓	~	X	X	X	✓

Legenda :

- ✓ Consigliato
- ~ Altra scelta possibile
- X Non compatibile

STRAUMANN – TISSUE LEVEL

	RN					WN				
	RN H4.0 / H6.0	RN ASC AH4.0 / AH5.5	RN Bridge/bar Conical	RN Bridge/bar Cylindrical	RN TAN Premill	WN H4.5 / H6.5	WN ASC AH4.5 / AH6.5	WN Bridge/bar Conical	WN Bridge/bar Conical	WN TAN Premill
ACI_IR	✓	✓	~	✓	✓					
ACI_IW						✓	✓	✓	✓	✓

Legenda :

✓ Consigliato

~ Altra scelta possibile

X Non compatibile

NOBEL BIO CARE

	Universal base CC NP 1.5	Universal base CC NP 3.0	CC NP Universal Base Non-Engaging Straight	CC NP Universal Base Non-Engaging Conical	CC NP Premill
NCCN_I.A1	~	X	X	X	✓
NCCN_I.A2	~	X	X	X	✓
NCCN_I.A3	X	✓	✓	X	✓
NCCN_I.A4	X	✓	✓	X	✓
NCCN_I.B1	~	X	X	X	✓
NCCN_I.B2	~	X	X	~	✓
NCCN_I.B3	X	✓	✓	X	✓
NCCN_I.B4	X	✓	✓	X	✓
NCCN_I.D1	✓	X	X	X	✓
NCCN_I.D2	✓	X	X	X	✓
NCCN_I.D3	X	✓	✓	X	✓
NCCN_I.D4	X	✓	✓	X	✓

	Universal base CC RP 1.5	Universal base CC RP 3.0	CC RP Universal Base Non-Engaging Straight	CC RP Universal Base Non-Engaging Conical	CC NP Premill
NCCR_I.B1	~	X	X	✓	✓
NCCR_I.B2	✓	X	X	✓	✓
NCCR_I.B3	X	~	✓	X	✓
NCCR_I.B4	X	✓	✓	X	✓
NCCR_I.C1	~	X	X	✓	✓
NCCR_I.C2	✓	X	X	✓	✓
NCCR_I.C3	~	~	✓	X	✓
NCCR_I.C4	X	✓	✓	X	✓
NCCR_I.D1	✓	X	X	~	✓
NCCR_I.D2	✓	X	X	~	✓
NCCR_I.D3	X	✓	✓	X	✓
NCCR_I.D4	X	✓	✓	X	✓

Legenda :

- ✓ Consigliato
- ~ Altra scelta possibile
- X Non compatibile

ZIMVIE

	Ti-Base GenTek TSV H4.0 / 4.7 Engaged / NonEngaged	AS-Ti-Base CH0.3 GenTek TSV H4.5 / 5.7 / 7.0 Engaged / NonEngaged	AS-Ti-Base CH1.3 GenTek TSV H4.5 / 5.7 / 7.0 Engaged / NonEngaged	AS-Ti-Base CH2.6 GenTek TSV H4.5 / 5.7 / 7.0 Engaged	Custom Abutment PSA GenTek TSV
ZSC35_I.A1	✓	✓	✗	✗	✓
ZSC35_I.A2	✓	✓	✗	✗	✓
ZSC35_I.A3	✓	✓	✗	✗	✓
ZSC35_I.A4	✓	✓	✗	✗	✓
ZSC35_I.B1	✓	✓	✗	✗	✓
ZSC35_I.B2	✓	✓	✗	✗	✓
ZSC35_I.B3	✓	✓	✗	✗	✓
ZSC35_I.B4	✓	✓	✗	✗	✓
ZSC35_I.C1	✓	✓	✗	✗	✓
ZSC35_I.C2	✓	✓	✗	✗	✓
ZSC35_I.C3	✓	✓	✗	✗	✓
ZSC35_I.C4	✓	✓	✗	✗	✓
ZSC35_I.D1	✓	✓	✗	✗	✓
ZSC35_I.D2	✓	✓	✗	✗	✓
ZSC35_I.D3	✓	✓	✗	✗	✓
ZSC35_I.D4	✓	✓	✗	✗	✓

	Ti-Base GenTek TSV H4.0 / 4.7 Engaged / NonEngaged	AS-Ti-Base CH0.3 GenTek TSV H4.5 / 5.7 / 7.0 Engaged / NonEngaged	AS-Ti-Base CH1.3 GenTek TSV H4.5 / 5.7 / 7.0 Engaged / NonEngaged	AS-Ti-Base CH2.6 GenTek TSV H4.5 / 5.7 / 7.0 Engaged	Custom Abutment PSA GenTek TSV
ZSC45_I.B1	✓	✓	✗	✗	✓
ZSC45_I.B2	✓	✓	✓	✗	✓
ZSC45_I.B3	✓	✓	✓	✗	✓
ZSC45_I.B4	✓	✓	✓	✗	✓
ZSC45_I.C1	✓	✓	✗	✗	✓
ZSC45_I.C2	✓	✓	✗	✗	✓
ZSC45_I.C3	✓	✓	✗	✗	✓
ZSC45_I.C4	✓	✓	✗	✗	✓
ZSC45_I.D1	✓	✓	✗	✗	✓
ZSC45_I.D2	✓	✓	✗	✗	✓
ZSC45_I.D3	✓	✓	✗	✗	✓
ZSC45_I.D4	✓	✓	✗	✗	✓

Legenda :

- ✓ Consigliato
- ~ Altra scelta possibile
- ✗ Non compatibile

BIOTECH

	Ti-base Ø4.0mm						Ti-base Ø5.0mm						Ti-base Ø6.5mm				Ti-base angolata						
	Custom Abutment BK-DI-36	KPCADCAM4007 /L	KPCADCAM4010 /L	KPCADCAM4020 /L	KPCADCAM4030 /L	KPCADCAM4040 /L	KPCADCAM4050 /L	KPCADCAM5007 /L	KPCADCAM5010 /L	KPCADCAM5020 /L	KPCADCAM5030 /L	KPCADCAM5040 /L	KPCADCAM5050 /L	KPCADCAM6520 /L	KPCADCAM6530 /L	KPCADCAM6540 /L	KPCADCAM6550 /L	KPCADCAM401VA	KPCADCAM402VA	KPCADCAM403VA	KPCADCAM501VA	KPCADCAM502VA	KPCADCAM503VA
BKO_I.A1	✓	✓	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	~	X	X	~	X	X
BKO_I.A3	✓	X	✓	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X	X
BKO_I.B1	✓	✓	X	X	X	X	X	✓	X	X	X	X	X	X	X	X	X	~	X	X	~	X	X
BKO_I.B3	✓	X	✓	X	X	X	X	X	X	✓	X	X	X	X	X	X	X	X	X	X	X	✓	X
BKO_I.C1	✓	✓	X	X	X	X	X	✓	X	X	X	X	X	X	X	X	X	~	X	X	~	X	X
BKO_I.C3	✓	X	✓	X	X	X	X	X	X	✓	X	X	X	X	~	X	X	X	X	X	X	✓	X

Legenda :

✓ Consigliato

~ Altra scelta possibile

X Non compatibile

GLOBAL D

		BaseTi Ø3.8mm				BaseTi Ø5.5mm			
	Custom Abutment S3DEL DLABPMVINK	DEVCI3.8H0.6	DEVCI3.8H2	DEVCI3.8H3	DEVCI3.8H4	DEVCI5.5H1.5	DEVCI5.5H2.2	DEVCI5.5H3	DEVCI5.5H4
GIK_I.A1	✓	✓	X	X	X	X	X	X	X
GIK_I.A3	✓	X	~	X	X	X	X	X	X
GIK_I.B1	✓	✓	X	X	X	~	X	X	X
GIK_I.B3	✓	X	X	X	X	X	✓	~	X
GIK_I.C1	✓	✓	X	X	X	X	X	X	X
GIK_I.C3	✓	X	X	X	X	X	✓	X	X

Legenda :

✓ Consigliato

~ Altra scelta possibile

X Non compatibile

MIS

	Ti-Base H=0.5, Anti-rotation (MD-TB001)	Ti-Base H=0.5, Free-rotation (MD-TB002)	Ti-Base H=1.5, Anti-rotation (MD-TBC15)	Ti-Base H=1.5, Free-rotation (MD-FTB15)	Ti-Base H=3.0, Anti-rotation (MD-TBC30)	Ti-Base H=3.0, Free-rotation (MD-FTB30)	EZ-Base SP H=1.5 (MD-EZB15)	Incisor Ti-Base H=0.5 C=6 (MD-IN060)	Ti-Blank SP
ZSC35_I.A1	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.A2	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.A3	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.A4	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.B1	✓	✓	✗	✗	✗	✗	✗	✗	✓
ZSC35_I.B2	✓	✓	✗	✗	✗	✗	✗	✗	✓
ZSC35_I.B3	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.B4	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.C1	✓	✓	✗	✗	✗	✗	✗	✗	✓
ZSC35_I.C2	✓	✓	✗	✗	✗	✗	✗	✗	✓
ZSC35_I.C3	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.C4	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.D1	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.D2	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.D3	✓	✓	✗	✗	✗	✗	✗	✓	✓
ZSC35_I.D4	✓	✓	✗	✗	✗	✗	✗	✓	✓

Legenda :

✓ Consigliato

~ Altra scelta possibile

✗ Non compatibile

	Ti-Base H=0.5, Anti-rotation (MW-TB001)	Ti-Base H=0.5, Free-rotation (MW-TB002)	Ti-Base H=1.5, Anti-rotation (MW-TBC15)	Ti-Base H=1.5, Free-rotation (MW-FTB15)	Ti-Base H=3.0, Anti-rotation (MW-TBC30)	Ti-Base H=3.0, Free-rotation (MW-FTB30)	EZ-Base WP H=1.5 (MW-EZB15)	Ti-Blank WP
ZSC45_I.B1	✓	✓	✗	✗	✗	✗	✗	✓
ZSC45_I.B2	✗	✗	✓	✓	✓	✓	✗	✓
ZSC45_I.B3	✗	✗	✗	✗	✓	✓	✗	✓
ZSC45_I.B4	✗	✗	✗	✗	✓	✓	✗	✓
ZSC45_I.C1	✓	✓	✗	✗	✗	✗	✗	✓
ZSC45_I.C2	✓	✓	✗	✗	✗	✗	✗	✓
ZSC45_I.C3	✗	✗	✓	✓	✗	✗	✓	✓
ZSC45_I.C4	✗	✗	✓	✓	✗	✗	✗	✓
ZSC45_I.D1	✗	✗	✗	✗	✗	✗	✗	✓
ZSC45_I.D2	✗	✗	✗	✗	✗	✗	✗	✓
ZSC45_I.D3	✗	✗	✗	✗	✗	✗	✗	✓
ZSC45_I.D4	✗	✗	✗	✗	✗	✗	✗	✓

ALPHABIO

	Ti-Base Engaged (Anti-R) GH 0.75 (4961) Ti-Base Non-Engaged (Rotational) GH 0.75 (4964)	Ti-Base Engaged (Anti-R) GH 1.5 (4962) Ti-Base Non-Engaged (Rotational) GH 1.5 (4965)	Ti-Base Engaged (Anti-R) GH2.5 (4963) Ti-Base Non-Engaged (Rotational) GH2.5 (4966)	Ti-Base Wide Engaged (Anti-R) GH 0.75 (5007) Ti-Base Wide Non-Engaged (Rotational) GH 0.75 (5008)	Blank PreFace
ZSC35_I.A1	✓	✗	✗	✗	✓
ZSC35_I.A2	✓	✗	✗	✗	✓
ZSC35_I.A3	✗	✓	✗	✗	✓
ZSC35_I.A4	✗	✓	✗	✗	✓
ZSC35_I.B1	✓	✗	✗	✓	✓
ZSC35_I.B2	✓	✗	✗	✓	✓
ZSC35_I.B3	✗	✓	✗	✗	✓
ZSC35_I.B4	✗	✓	✗	✗	✓
ZSC35_I.C1	✗	✗	✗	✓	✓
ZSC35_I.C2	✓	✗	✗	✓	✓
ZSC35_I.C3	✗	✓	✗	✗	✓
ZSC35_I.C4	✗	✓	✗	✗	✓
ZSC35_I.D1	✓	✗	✗	✗	✓
ZSC35_I.D2	✓	✗	✗	✗	✓
ZSC35_I.D3	✗	✓	✗	✗	✓
ZSC35_I.D4	✗	✓	✗	✗	✓

Legenda :

✓ Consigliato

~ Altra scelta possibile

✗ Non compatibile

ANTHOGYR

	Flexibase BL single 4.0 H1.5 (OPFLEX413)	Flexibase BL single 4.0 H2.5 (OPFLEX423)	Flexibase BL single 4.0 H3.5 (OPFLEX433)	Flexibase BL single 5.0 H1.5 (OPFLEX513)	Flexibase BL single 5.0 H2.5 (OPFLEX523)	Flexibase BL single 5.0 H3.5 (OPFLEX533)	PreFace
AAX_I.A1	X	X	X	X	X	X	✓
AAX_I.A3	✓	X	X	X	X	X	✓
AAX_I.B1	X	X	X	X	X	X	✓
AAX_I.C3	✓	X	X	X	✓	X	✓
AAX_I.C1	X	X	X	X	X	X	✓
AAX_I.C3	✓	X	X	X	✓	X	✓

Legenda :

✓ Consigliato

~ Altra scelta possibile

X Non compatibile

BIOHORIZONS

	Hexed (PYHYB)	Non-Hexed (PYHYBN)	Hexed 2mm (PYHYB2)	Laser Lock (PYTBL)	Ti	Ti Laser Lock
ZSC35_I.A1	✓	✓	✗	✗	✓	✗
ZSC35_I.A2	✓	✓	✗	✗	✓	✗
ZSC35_I.A3	✓	✓	✗	✓	✓	✗
ZSC35_I.A4	✓	✓	✗	✓	✓	✗
ZSC35_I.B1	✓	✓	✗	✗	✓	✓
ZSC35_I.B2	✓	✓	✗	✗	✓	✓
ZSC35_I.B3	✓	✓	✗	✓	✓	✗
ZSC35_I.B4	✓	✓	✗	✓	✓	✗
ZSC35_I.C1	✓	✓	✗	✗	✓	✗
ZSC35_I.C2	✓	✓	✗	✗	✓	✓
ZSC35_I.C3	✓	✓	✗	✓	✓	✗
ZSC35_I.C4	✓	✓	✗	✓	✓	✗
ZSC35_I.D1	✓	✓	✗	✗	✓	✗
ZSC35_I.D2	✓	✓	✗	✗	✓	✗
ZSC35_I.D3	✓	✓	✗	✓	✓	✗
ZSC35_I.D4	✓	✓	✗	✓	✓	✗

	Hexed (PGHYB)	Non-Hexed (PGHYBN)	Hexed 2mm (PGHYB2)	Laser Lock (PGTBL)	Ti	Ti Laser Lock
ZSC45_I.B1	✓	✓	✗	✗	✓	✗
ZSC45_I.B2	✓	✓	✗	✓	✓	✗
ZSC45_I.B3	✓	✓	✓	✓	✓	✗
ZSC45_I.B4	✓	✓	✓	✓	✓	✗
ZSC45_I.C1	✓	✓	✗	✗	✓	✓
ZSC45_I.C2	✓	✓	✗	✗	✓	✗
ZSC45_I.C3	✓	✓	✓	✓	✓	✗
ZSC45_I.C4	✓	✓	✓	✓	✓	✗
ZSC45_I.D1	✓	✓	✗	✗	✓	✗
ZSC45_I.D2	✓	✓	✗	✗	✓	✗
ZSC45_I.D3	✓	✓	✗	✓	✓	✗
ZSC45_I.D4	✓	✓	✗	✓	✓	✗

Legenda :

- ✓ Consigliato
- ~ Altra scelta possibile
- ✗ Non compatibile

ORDINA IL TUO COFANETTO !

Questo kit vi mette a disposizione l'insieme dei prodotti necessari alla realizzazione dei vostri primi casi clinici :

- Un assortimento di 6 o 9 Profile Designer iPhysio® sterili e monouso (uno per ogni forma A, B, C, D e altezza)
- 1 kit prova iPhysio® Profile Designer con iPhysio® di prova riutilizzabili (uno per ogni forma e altezza)
- 3 monconi provvisori iPhysio®



Compatibilità	Codice
Anthogyr® - Axiom® BL REG, PX e X3	CDIAAX
Biotech Dental® - Kontakt®, Kontakt S®	CDIBKO
GlobalD® - In-Kone® Universal e Primo	CDIGIK
Nobel Biocare® - NobelActive®, NobelReplace® CC e NobelParallel® CC - Implant Direct® - InterActive®	CDINCC
Straumann® - Bone Level (BL) e Bone Level Tapered (BLT)	CDISBL
ZimVie® - Tapered Screw-Vent®, Trabecular Metal™, TSX™ / MIS® - SEVEN® e M4® - Implant Direct® - Legacy® / BioHorizons® - Internal, Tapered Internal, Tapered Plus	CDIZSC
Alpha-Bio Tec® - NeO, SPI, I.C.E., DFI, ATID / A.B. Dental® - I2, I22, I5, I55, I10	CDIZSC35
LYRA ETK® Naturactis, Naturall+, Natea+, Astra Tech® - OsseoSpeed™ TX Ocean/Aqua - Easy Implant® - Master-C, Master-S Ocean	CDIE3N

LYRA-ETK
DIGITAL DENTAL PROTOCOLS

Simple . Safe . Accessible